



Project No. (Item)	004635
	QA
Inspection done by & Date	B.Roosdorp 11-april-2024
System Information	
Manufacturer and model	GE Brightspeed 16
Upgraded System? <ul style="list-style-type: none"> • From which one? • Year of the upgrade? 	Yes <input type="checkbox"/> No <input checked="" type="checkbox"/>
System serial no.	671917CT01
System YOM	2011
Number of slices	Water
Air or water cooling?	Air
Gantry & Detector Information	
Gantry YOM	2011
Gantry model no.	5191003
Gantry serial no.	267068HM5
Gantry count (rotations)	6761695
Detector model no.	5488088 - PDAS16
Detector serial no.	267104HM8
Detector YOM	January 2011
Tube Information	
Tube Type	Solarix 350 Brightspeed
Tube Housing model no.	2199077-7
Tube Housing serial no.	180924BC3
Tube Insert model no.	46-309300G3
Tube Insert serial no.	165516BI6
Tube YOM	2021
Tube installation date	14 th June 2021
Tube count(s) (mAs, secs, count etc.)	Tube mAs: 25 801 417.5 Scan Seconds: 176 231.7
Patient Table Information	
Table model no.	5182488
Table serial no.	266595HM8

Table YOM	2011
Head rest present?	Yes v NO <input type="checkbox"/>
Positioning pads incl.?	Yes v NO <input type="checkbox"/>
PDU Information	
PDU/transformer model no.	2326492-3
PDU/transformer serial no.	266879HM6
PDU/transformer YOM	2011
WORKSTATION INFORMATION (Main-Host)	
Workstation model no.	5330382
Workstation serial no.	267038HM8
GOC Version	GEHC/CTT Linux 6.3.11
Workstation software version	11BW46.3_SP1-9-1 11BW46.3_SP1-9-1 Patch
Workstation software options	Helical-120-Sec Lite-Power350 Smart Prep Patient-16-slice AutomA Direct-3D Direct-MPR Helical Tilt 3000 Image Series Data Export CopyComposer NeuroFilter VolumeViewer DentaScan Connect Pro
ACCESSORIES, MANUAL HARD-COPIES AND SOFTWARE DISCS	
Software	Yes v NO <input type="checkbox"/>
Manuals	Yes v NO <input type="checkbox"/>
MAINTENANCE	
Any major part replacements/service done?	1. Last service, software reload. 23 may performance test planned
Is system logbook present?	Yes v

COMMENTS

Functional condition	Normal working condition
Cosmetic condition	Good

PHANTOMS

Model name	Model no.	Serial no.
QA phantom	2206351	1000183
QA phantom	2144715	1006405
Body phantom - white	Na	Na

Q&A TESTING

	Tests Names Performed	RESULT OF TESTS
		QA
1	Warm up	Pass
2	Air calibration	Pass
3	General service scan water and performance	Pass